

Invest Ophthalmol Vis Sci 2009;50: E-Abstract 1232.© 2009 [ARVO](#)

1232

Quality of the Tear Film Measured With an Objective Optical Method

G. M. Perez¹, M. Fernandez¹, S. Mirabet², F. Diaz-Douton³,
J. Pujol³, J. M. Marin² and P. Artal¹¹Laboratorio de Optica, Universidad de Murcia, Murcia, Spain²Hospital Virgen de la Arrixaca, Servicio de Oftalmologia, Murcia, Spain³CD6, Universidad Politecnica de Cataluña, Tarrasa, Spain**Commercial Relationships:** G.M. Perez, None; M. Fernandez,

None; S. Mirabet, None; F. Diaz-Douton, Visiometrics, C; J. Pujol, Visiometrics, I; Visiometrics, P;

J.M. Marin, None; P. Artal, Visiometrics, F; Visiometrics, I; Visiometrics, P.

Support: "Ministerio de Educación y Ciencia", Spain (grant n° FIS2007-64765) and "Fundación Séneca", Murcia, Spain (grant 04524/GERM/06) and Visiometrics SL.

Abstract

Purpose: The quality of the tear film affects severely the quality of vision. The purpose of this study is to develop a novel optical non-invasive approach to characterize tear film quality. It is based in the analysis of the ocular scattering dynamics measured with a double-pass instrument.**Methods:** The procedure consisted in the dynamic recording of double-pass (DP) images (Santamaria et al., JOSAA, 1987) during unforced tear film break-up. Series of images (every second) were recorded in the eyes of a group of both healthy and dry-eye patients with a DP-based instrument (OQAS, Visiometrics, Spain). An unstable tear film would produce an elevation of the ocular scattering and in consequence a degraded retinal image. The relative dynamic changes of scatter would be related with modifications in the tear film quality. The dynamics of ocular scattering was evaluated with an Objective Scatter Index (OSI) calculated for each individual DP image. This parameter quantifies the level of scatter and its variations with time will indicate the impact of tear film deterioration.**Results:** In healthy eyes with good quality tear film, the analysis of the series of DP images showed that normal tear film break-up process involves minor fluctuations in the OSI value (0.5 ± 0.2), whereas an abnormally accelerated tear film break-up tends to increase slightly this value (0.9 ± 0.3). Patients with diagnosed dry eyes showed both an increased average OSI value and higher fluctuations (3 ± 2). The*This Article**Services*

- ▶ [Email this article to a friend](#)
- ▶ [Similar articles in this journal](#)
- ▶ [Alert me to new issues of the journal](#)
- ▶ [Download to citation manager](#)

Citing Articles

- ▶ [Citing Articles via Google Scholar](#)

Google Scholar

- ▶ [Articles by Perez, G. M.](#)
- ▶ [Articles by Artal, P.](#)
- ▶ [Search for Related Content](#)

PubMed

- ▶ [Articles by Perez, G. M.](#)
- ▶ [Articles by Artal, P.](#)

quality of the tear film can be determined and graded by the average value and standard deviation of the scatter parameter.

Conclusions: A new robust and objective optical method to quantify the quality and stability of the tear film has been developed. It is based on measuring the induced changes in the scattering that affect the retinal image. This technique may be useful to detect and follow-up tear-film related patient's complaints.

Keywords: cornea: tears/tear film/dry eye • optical properties • image processing



© 2009, The Association for Research in Vision and Ophthalmology, Inc., all rights reserved. Permission to republish any abstract or part of an abstract in any form must be obtained in writing from the ARVO Office prior to publication.

[HOME](#) [HELP](#) [FEEDBACK](#) [SUBSCRIPTIONS](#) [ARCHIVE](#) [SEARCH](#)